EXHIBIT B-4

From: Jaccard, Jamil [JNJCH] <jjaccard@its.jnj.com>

Sent: Tue, Aug. 17, 2010 14:52:06 GMT

To: Heron, Claire [JNJCH] < CHERON@its.jnj.com>

Cc: Poulot, Stéphane [JNJCH] <spoulot@its.jnj.com>; Hostache, Karine [JNJCH] < khostach@its.jnj.com>;

EI Fettouhi, Atimad [JNJCH] <aelfetto@ITS.JNlcom>

Subject: Particles in production

Hello Claire,

Pursuant to your request at the TAM, please find the presentation attached:

I am at your disposal for any questions.

Best Regards,


Jamil Jaccard QA Engineer ETHICON SÀRL A Johnson and Johnson Company Puits Godet, 20 2000 Neuchâtel - Switzerland Tel: (+41) 32 934 86 96

Fax: (+41) 32 934 86 90 Fax: (+41) 32 934 86 10 Email: jjaccard@its.jnj.com

<u>Placeholder</u> Document produced in native format

Case 2:12-md-02327	Document 1750-4	Filed 10/27/15	Page 4 of 12 PageID #: 21829
EVOLUTION OF NCS		ATED TO FORE IU in 2010	IGN MATTER FOR SWC AND


EVOLUTION OF NCRS RELATED TO FOREIGN MATTER SINCE 2009 FOR NEUCHÂTEL:



[Linéaire = Linear]

Type of Foreign Matter at the source of NCRs, SWC + EWHU in 2010:


Type of particles related to 2010 NCRs SWC + EWHU



Note: Metallic are related to the TVT needle problem

2.1) ONLY FOR EWHU:


Type of particles related to 2010 NCRs EWHU



Note: Metallic are related to the TVT needle problem

2.2) ONLY FOR SWC US

Type of particles related to 2010 NCRs SWC US




2.3) SWC EMEA NO NCRs in 2010

3.) **EDOC0161** EWHU 2010:

(DOES NOT REVEAL THE TYPE OF FOREIGN MATTER)

EDOC0161: Evolution of foreign matter rejects in 2010


PROLIFT TVTO TVT P1990X



3.2) EDOC0161 SWC

(SWC US WITH DELAY SINCE APRIL)

EDOC0161: Evolution of foreign matter rejects in 2010



(SWC US Delayed since April)

CONCLUSIONS:

- A big spike in NCRs in March showed the indoor Foreign Matter problem.
- The main source is human for all types of NCRs (In process/complaint/Raw material)
- Indoor component rejects (according to EDOC161) are stable in 2010.
- According to the number of NCRs and the EDOC0161, SWC US is more sensitive to particles (or dirt) than SWC EMEA.

CERTIFICATE OF TRANSLATION

STATE OF NEW YORK

8

COUNTY OF NEW YORK §

Steven J. Albert, being duly sworn, deposes and says:

- 1. I am employed as an Account Executive for LanguageWorks, a professional translation firm providing translation services since 1993. I have personal knowledge of all the facts stated herein.
- 2. Attached to this Affidavit is a translation from French into English made by LanguageWorks of **ETH.MESH.13907354-13907355.**
- 3. Language Works only utilizes the services of language professionals who have passed our stringent testing process, which involves a subject matter and language-specific test translation that is reviewed by trusted linguists. Less than 6% of applicants pass. In addition, our language professionals must have a minimum of 5 years' professional translation experience or an applicable degree from an accredited institution. In addition, for all translations, including this one, Language Works uses three qualified linguists: a professional translator; an editor to double-check for style and content; and a proofreader to review all text for completeness. Translations made through our processes set forth above are regularly accepted into evidence in New York State courts.
- 4. On the basis of the foregoing process and quality control procedures, I certify that the attached is an accurate translation of the original.

Steven J. Albert

SUBSCRIBED AND SWORN TO before me this twenty-first day of October, 2015.

Notary Public

DAWN IN JORDAN
NOTARY PUBLIC STATE OF NEW YORK
NO. 01J06328681

Qualified in Kings County
My Commission Expires August 03, 2019